

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: March 4, 1992
 Date Submitted: February 26, 1992
 Project: 7238, Metro Self Monitor, PO #M23390

RESULTS OF ANALYSES OF THE WATER SAMPLE FOR SELECTED METALS BY ICP (6010) Results Reported as mg/L (ppm)

<u>Sample #</u>	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M23390	1.3	0.81	0.78	0.07
<u>Quality Assurance</u>				
Method Blank	<0.02	<0.02	<0.02	<0.02
M23390 (Duplicate)	1.2	0.29	0.75	0.07
M23390 (Matrix Spike) Percent Recovery	97%	97%	97%	91%
M23390 (Matrix Spike Duplicate) Percent Recovery	97%	97%	97%	91%
Spike Level	5	5	10	5

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: March 4, 1992
Date Submitted: February 26, 1992
Project: 7238, Metro Self Monitor, PO #M23390

**RESULTS OF ANALYSES OF THE WATER SAMPLE
FOR SELECTED METALS BY ICP (6010)
Results Reported as mg/L (ppm)**

<u>Sample #</u>	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M23390	1.3	0.81	0.78	0.07
<u>Quality Assurance</u>				
Method Blank	<0.02	<0.02	<0.02	<0.02
M23390 (Duplicate)	1.2	0.29	0.75	0.07
M23390 (Matrix Spike) Percent Recovery	97%	97%	97%	91%
M23390 (Matrix Spike Duplicate) Percent Recovery	97%	97%	97%	91%
Spike Level	5	5	10	5

SAMPLE CHAIN OF CUSTODY

SITE NO.

PURCHASE ORDER #

7238

METRO SELF. MONITOR

M 23390

PROJECT LOCATION

REMARKS

SAMPLE DISPOSAL INFORM.

- ☐ Dispose after 30 days
☒ Return Samples
☐ Call for Instructions

Date/Time
Sampled

Type of Sample

of
Jars

Lab
Sample #

**Analyses
Requested**

m 23390

Water

/

27324

Cr, Cu, Ni, Zn

PRINT NAME _____

COMPANY

Date _____

Time

Relinquished by:

Received by:

Relinquished by:

Received by:

AKC-0014101

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Andrew John Friedman
James E. Bruya, Ph.D.
(206) 285-8282

3008-B 16th Avenue West
Seattle, WA 98119
FAX: (206) 283-5044

March 4, 1992

Greg Speer, Project Leader
Alaskan Copper
628 South Hanford
Seattle, WA 98134

Dear Mr. Speer:

Enclosed are the results of the analyses of the sample submitted on February 26, 1992 from Project 7238, Metro Self Monitor, PO #M23390.

We appreciate this opportunity to be of service to you on this project. If you have any questions regarding this material, or if you just want to discuss any aspect of your projects, please do not hesitate to contact me.

Sincerely,



Mark Z Perin, Chemist

MZP/dp

Enclosures